

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100387

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 04/27/2012

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.74: 0.079  
LOT #: 02511080A3  
EXPIRATION: 02/01/2013

BLANK TEST	0.000	18:54
INTERNAL STANDARD	VERIFIED	18:54
EXTERNAL STANDARD	0.078	18:55
BLANK TEST	0.000	18:56
EXTERNAL STANDARD	0.078	18:56
BLANK TEST	0.000	18:57
EXTERNAL STANDARD	0.078	18:57
BLANK TEST	0.000	18:58
EXTERNAL STANDARD	0.078	18:58
BLANK TEST	0.000	18:59
EXTERNAL STANDARD	0.078	18:59
BLANK TEST	0.000	19:00

Average = 0.0780  
Std Dev = 0.0000

## Diagnostic Check

### VERSIONS

DMT: 1.05  
PIC: 3.02  
Modem: 2.0  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.6°C PASSED  
Breath Tube = 44.7°C PASSED

### PUMP INFO

Flow Rate = 5.083 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED


### FILTER INFO

Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

INTERNAL STANDARD PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

7/3/2012

Subscribed and sworn before me this 3rd day of July, 20 12

  
Nikki Roth, Notary Public  
My Commission Expires With Office



  
BMB  
6/24/12