VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument
State of Alaska

Serial #: 100414 — ***Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*** — Date:10/10/2011

External Standard Test Values

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.34: 0.078
LOT #: 34409080A1
EXPIRATION: 01/01/2012

BLANK TEST	0.000	00:08
INTERNAL STANDARD	VERIFIED	00:08
EXTERNAL STANDARD	0.079	00:08
BLANK TEST	0.000	00:09
EXTERNAL STANDARD	0.079	00:09
BLANK TEST	0.000	00:10
EXTERNAL STANDARD	0.079	00:10
BLANK TEST	0.000	00:11
EXTERNAL STANDARD	0.079	00:12
BLANK TEST	0.000	00:12
EXTERNAL STANDARD	0.080	00:13
BLANK TEST	0.000	00:14

Average = 0.0792
Std Dev = 0.0004

Diagnostic Check

VERSIONS
DMT: 1.01
PIC: 3.02
Modem: 2.0
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C	PASSED
Breath Tube = 45.8°C	PASSED

PUMP INFO

Flow Rate = 5.337 L/M	PASSED

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz 11/4/11

Nita J.Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 4 day of Nov, 2011

Nikki Roth

Nikki Roth, Notary Public
My Commission Expires With Office

NIKKI ROTH
My Commission Expires
NOTARY PUBLIC
With Office
STATE OF ALASKA

BMB
10/28/11