

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130112</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Bruce Smith Jr</u>	ID# <u>3522</u> ✓ Date <u>3-1-11</u> ✓
A	
Agency <u>Wrangell Police Department</u>	Phone # <u>(907) 874-3304</u>
Instrument Location <u>Wrangell Police Department, 431 Zimovia Highway</u>	
B	
Alco S/N <u>X301542</u> ✓	Target Value <u>.077</u> ✓ High Pressure <u>750</u>
Alco Test Values	<u>.085</u> ✓ 1st Alco
	<u>.083</u> ✓ 2nd Alco
Signature <u>B Smith</u>	

(OVER)

BMB
3/1/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 31 day of MAR, 2011.



[Signature]
 (Notary Seal Stamp)
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130112

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130112

MARCH 01, 2011

OPERATOR'S NAME:

SMITH JR, BRUCE W

OPERATOR'S NUMBER: 3522

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/NI :

UOC

O.L. #: 0123456

DEPT/AGENCY: MR61

CASE/REPORT: 11030000

TEST TYPE: U

ALCO TARGET VALUE: .077

ALCO S/N: X301542

— BREATH ANALYSIS —

.077 ADJUSTED FOR 29.99 in
ALCO TARGET .077 04:30
BLANK TEST .000 04:31
INTERNAL STANDARD VERIFIED 04:31
ALCO TV 29.99 in .085 04:31
BLANK TEST .000 04:32
SUBJECT SAMPLE .000 04:33
BLANK TEST .000 04:33
ALCO TV 29.99 in .083 04:34
BLANK TEST .000 04:35

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130112

MARCH 01, 2011

TIME 04:28

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 42c
BAROMETER: 29.99 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNPQRSTUWXYZ[\]^_`abcdefgijklmnop
qrsstuvwxyz{|}~||