

VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument
State of Alaska*

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130112</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Bruce Smith Jr</u>	ID# <u>3522</u> Date <u>4-19-11</u>
A Agency <u>Wrangell Police Department</u> Phone # <u>(907) 874-3304</u>	
Instrument Location <u>Wrangell Police Department, 431 Zimovia Highway</u>	
B Alco S/N <u>X301542</u> Target Value <u>.077</u> High Pressure <u>450</u>	
Alco Test Values	<u>.084</u> <u>.084</u> 1st Alco 2nd Alco
Signature <u></u>	


(OVER)

*BMB
4/27/11*

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 31 day of MAY, 2011.

 (Notary Seal Stamp)
 Daniel L. Nicholas Kyle Klopfer
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130112

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130112

APRIL 19, 2011

OPERATOR'S NAME:
SMITH JR, BRUCE W
OPERATOR'S NUMBER: 3522
SUBJECT'S LAST NAME:
UOC
SUBJECT'S FIRST NAME/MI :
UOC

O.L. #: 0123456
DEPT/AGENCY: MRS1
CASE/REPORT: 11040000
TEST TYPE: U
ALCO TARGET VALUE: .077
ALCO S/N: X301542

— BREATH ANALYSIS —

.077 ADJUSTED FOR 30.06 in
ALCO TARGET .077 17:19
BLANK TEST .000 17:20
INTERNAL STANDARD VERIFIED 17:20
ALCO TU 30.06 in .004 17:20
BLANK TEST .000 17:21
SUBJECT SAMPLE .000 17:21
BLANK TEST .000 17:22
ALCO TU 30.06 in .004 17:23
BLANK TEST .000 17:23

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130112

APRIL 19, 2011
TIME 17:24

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 43c
BAROMETER: 30.06 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; (=) ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢

