

FEB 08 2008

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument*

*State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130112</u>
Name	<u>Bruce Smith Jr</u>	ID#	<u>3522</u> Date <u>2-4-08</u>
A Agency		<u>Wrangell Police Department</u>	Phone # <u>(907) 874-3304</u>
Instrument Location <u>Wrangell Police Department, 431 Zimovia Highway</u>			
B Alco S/N		<u>X 301548<sup>✓</sup></u>	Target Value <u>.080<sup>✓</sup></u> High Pressure <u>550</u>
Alco Test Values		<u>.082</u>	<u>.082</u>
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature		<u><i>[Signature]</i></u> <i>CSB 2/21/08</i>	

(OVER)

Under the Alaska Rules of Evidence, I certify that;

(Do Not write in the area below)

(a) I am Jeanne Swartz, Criminalist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Bruce W Smith, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed

*[Signature]*

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

02/22/08  
Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130112

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130112

FEBRUARY 04, 2008

OPERATOR'S NAME:

SMITH JR, BRUCE W

OPERATOR'S NUMBER: 3522

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/NI :

CAL

D.L. #: 012345

DEPT/AGENCY: MRG1

CASE/REPORT: 06020000

TEST TYPE: U

ALCO TARGET VALUE: .000

ALCO S/N: X301548

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.85 in  
ALCO TARGET .079 14:46  
BLANK TEST .000 14:47  
INTERNAL STANDARD VERIFIED 14:47  
ALCO TV 29.85 in .002 14:47  
BLANK TEST .000 14:48  
SUBJECT SAMPLE .000 14:49  
BLANK TEST .000 14:49  
ALCO TV 29.85 in .002 14:50  
BLANK TEST .000 14:51

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130112

FEBRUARY 04, 2008

TIME 14:45

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 43c  
BAROMETER: 29.85 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF  
GHIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~■