

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130112 ✓
Name		Bruce Smith Jr	ID# 3522 Date 2-4-09 ✓
A	Agency	Wrangell Police Department	Phone # (907) 874-3304
Instrument Location		Wrangell Police Department, 431 Zimovia Highway ✓	
Alco S/N		X301583 ✓	Target Value .080 ✓ High Pressure 500
B	Alco Test Values	.085 ✓ 1 st Alco	.084 ✓ 2 nd Alco
Signature			COB 2/18/09


(OVER)

Under the Alaska Rules of Evidence, I certify that (Do Not write in the area below)


(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Bruce W Smith, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,



Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

2/18/09

Date

BT9 06/05 FEB 09 2009

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130112 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130112

FEBRUARY 04, 2009

OPERATOR'S NAME: *CAS 2/18/09*
SMITH JR, BRUCE W
OPERATOR'S NUMBER: 3522
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
CAL

O.L. #: 012345
DEPT/AGENCY: WSG1
CASE/REPORT: 09020000
TEST TYPE: U
ALCO TARGET VALUE: .000 ✓
ALCO S/N: X301583 ✓

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.81 in		
ALCO TARGET	.079	21:41
BLANK TEST	.000	21:42
INTERNAL STANDARD	VERIFIED	21:42
ALCO TV 29.81 in	.085 ✓	21:43
BLANK TEST	.000	21:43
SUBJECT SAMPLE <i>CAS 2/18/09</i>	.000	21:44
BLANK TEST	.000	21:45
ALCO TV 29.81 in	.084 ✓	21:45
BLANK TEST	.000	21:46

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130112

FEBRUARY 04, 2009 ✓
TIME 21:39

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 29.81 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; (=) ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~