


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130112 ✓			
Name	Bruce Smith Jr	ID#	3522	Date	3-31-08 ✓	
A	Agency	Wrangell Police Department	Phone #	(907) 874-3304		
Instrument Location						Wrangell Police Department, 431 Zimovia Highway
B						
Alco S/N	X301548 ✓	Target Value	.080 ✓	High Pressure	300	
Alco Test Values	.085		.084			
	1 st Alco		2 nd Alco			
Signature					 4/8/08	
(OVER)						

Under the Alaska Rules of Evidence, I certify that (Print name in the area below)


(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Bruce W Smith, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,



Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

4/12/08
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130112

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130112

MARCH 31, 2008

COB 4/8/08
OPERATOR'S NAME:
SMITH JR, BRUCE W
OPERATOR'S NUMBER: 3522
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/NI :
CAL
O.L. #: 012345
DEPT/AGENCY: WR61
CASE/REPORT: 00030000
TEST TYPE: U
ALCO TARGET VALUE: .000
ALCO S/N: X301548

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.69 in
ALCO TARGET .002 13:38
BLANK TEST .000 13:39
INTERNAL STANDARD VERIFIED 13:39
ALCO TU 30.69 in .005 13:39
BLANK TEST .000 13:40
SUBJECT SAMPLE .000 *COB 4/8/08* 13:40
BLANK TEST .000 *4/8/08* 13:41
ALCO TU 30.69 in .004 13:42
BLANK TEST .000 13:42

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130112

MARCH 31, 2008
TIME 13:36

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 42c
BAROMETER: 30.69 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!""%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNO PQRSTU VWXYZ[\]^_`abc defghijklmno
pqr stuvwxyz{|}~■