

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

JUN 16 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130112</u> ✓	
Name <u>Bruce Smith Jr</u>		ID# <u>3522</u> Date <u>6-8-09</u> ✓	
A	Agency <u>Wrangell Police Department</u>	Phone # <u>(907) 874-3304</u>	
Instrument Location <u>Wrangell Police Department, 431 Zimovia Highway</u> ✓			
Alco S/N <u>X301583</u> ✓ Target Value <u>.080</u> ✓ High Pressure <u>250</u>			
B	Alco Test Values	<u>.084</u> ✓ 1 st Alco	<u>.084</u> ✓ 2 nd Alco
Signature <u>[Signature]</u> ✓			
(OVER)		<u>[Signature]</u> 6/16/09	

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 24th day of June 2009.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130112

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130112

JUNE 08, 2009 ✓

*Bmb
6/16/09*

OPERATOR'S NAME:
SMITH JR BRUCE W
OPERATOR'S NUMBER: 3522
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
CAL

O.L. #: 012345
DEPT/AGENCY: WRG1
CASE/REPORT: 09060000
TEST TYPE: V
ALCO TARGET VALUE: .000 ✓
ALCO S/N: X301583 ✓

— BREATH ANALYSIS —

.000 ADJUSTED FOR 30.24 in ✓
ALCO TARGET .000 07:45
BLANK TEST .000 07:46
INTERNAL STANDARD VERIFIED 07:46
ALCO TV 30.24 in .004 ✓ 07:46
BLANK TEST .000 ✓ 07:47
SUBJECT SAMPLE *Bmb* .000 07:47
BLANK TEST .000 07:48
ALCO TV 30.24 in *6/16/09* .004 ✓ 07:48
BLANK TEST .000 07:49

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130112

JUNE 08, 2009 ✓
TIME 07:41

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 43c
BAROMETER: 30.24 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLHINOPQRSTUWXYZ\]^_`'abcde fghijklmno
pqrstuvwxyz{|}~■