

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

NOV 02 2009

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130112 ✓
Name		Bruce Smith Jr	ID# 3522 Date 10-27-09 ✓
A	Agency	Wrangell Police Department	Phone # (907) 874-3304
Instrument Location Wrangell Police Department, 431 Ylmovia Highway ✓			
Alco S/N		X 172933	Target Value .080 High Pressure 600
B	Alco Test Values	.087 ✓ 1 st Alco	.086 ✓ 2 nd Alco
Signature 			
(OVER)			

BMS
11/14/09


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 30 day of Dec., 2009.


(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130112

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130112

OCTOBER 27, 2009

OPERATOR'S NAME:
SMITH JR, BRUCE W
OPERATOR'S NUMBER: 3322
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :

CAL
O.L. #: 0123456
DEPT/AGENCY: MRS1
CASE/REPORT: 09100000
TEST TYPE: U
ALCO TARGET VALUE: .000
ALCO S/N: X172933

BREATH ANALYSIS

ALCO TARGET	.001	16:11
BLANK TEST	.000	16:12
INTERNAL STANDARD	VERIFIED	16:12
ALCO TV 30.51 in	.007	16:12
BLANK TEST	.000	16:13
SUBJECT SAMPLE	.000	16:13
BLANK TEST	.000	16:14
ALCO TV 30.51 in	.006	16:14
BLANK TEST	.000	16:15

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130112

OCTOBER 27, 2009
TIME 16:09

DIAGNOSTIC CHECK

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 40c
BREATH TUBE: 41c
BAROMETER: 30.51 in
FLOW DETECTOR: OKAY

DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ\]^_`*abcdefghijklmnopqr
stuvwxyz{|}~@