

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130112 ✓			
Name	Bruce Smith Jr	ID#	3522	Date	12-8-08 ✓	
A	Agency	Wrangell Police Department	Phone #	(907) 874-3304		
Instrument Location		Wrangell Police Department, 431 Zimovia Highway ✓				
Alco S/N		X 301583 ✓	Target Value	.080 ✓	High Pressure	650
B	Alco Test Values		.085 ✓ 1 st Alco	.084 ✓ 2 nd Alco		
Signature		 ✓			COB 12/8/08	
(OVER)						

Under the Alaska Rules of Evidence, I certify that (I am the only one who wrote in the area below)


(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Bruce W Smith, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,



Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

12/20/08
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130112

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130112

DECEMBER 08, 2008

OPERATOR'S NAME: SMITH JR, BRUCE W
OPERATOR'S NUMBER: 3522
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
CAL

O.L. #: 0123456
DEPT/AGENCY: MRG1
CASE/REPORT: 08120000
TEST TYPE: U
ALCO TARGET VALUE: .080
ALCO S/N: X301583

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 30.17 in
ALCO TARGET .080 15:19
BLANK TEST .080 15:20
INTERNAL STANDARD VERIFIED 15:20
ALCO TV 30.17 in .085 ✓ 15:20
BLANK TEST .080 15:21
SUBJECT SAMPLE .080 15:21
BLANK TEST .080 15:22
ALCO TV 30.17 in .084 ✓ 15:22
BLANK TEST .080 15:23

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130112

DECEMBER 08, 2008
TIME 15:17

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 30.17 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z I \] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~