

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DEC 21 2009

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130112

Name Bruce Smith Jr ID# 3522 Date 12-18-09

A Agency Wrangell Police Department Phone # (907) 874-3304

Instrument Location Wrangell Police Department, 431 Zimovia Highway

B Alco S/N X172933 Target Value .080 High Pressure 500

Alco Test Values .084 .084
1st Alco 2nd Alco

Signature [Signature]

(OVER)

BMB
1/5/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 14th day of January, 2010
[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130112

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130112

DECEMBER 18, 2009

OPERATOR'S NAME:
SMITH JR, BRUCE W
OPERATOR'S NUMBER: 3522
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/NI :
CAL

O.L. #: 0123456
DEPT/AGENCY: WR61
CASE/REPORT: 09120000
TEST TYPE: U
ALCO TARGET VALUE: .000
ALCO S/N: X172933

BREATH ANALYSIS

.000 ADJUSTED FOR 29.54 in
ALCO TARGET .079 00:35
BLANK TEST .000 00:36
INTERNAL STANDARD VERIFIED 00:36
ALCO TV 29.52 in .084 00:37
BLANK TEST .000 00:37
SUBJECT SAMPLE .000 00:38
BLANK TEST .000 00:39
ALCO TV 29.54 in .084 00:39
BLANK TEST .000 00:40

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130112

DECEMBER 18, 2009
TIME 00:41

DIAGNOSTIC CHECK

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 29.52 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!*"#%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNPOQRSTUVWXYZ0123456789:;
abcdefghijklmnopqrstuvwxyz{|}~