

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

FEB 10 2010

Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130112</u>	
Name <u>Bruce Smith Jr</u>	ID# <u>3522</u> Date <u>2-9-10</u>
A Agency <u>Wrangell Police Department</u> Phone # <u>(907) 874-3304</u>	
Instrument Location <u>Wrangell Police Department, 431 Zimovia Highway</u>	
B Alco S/N <u>77199</u> Target Value <u>.083</u> High Pressure <u>1000</u>	
Alco Test Values	<u>.088</u> <u>.088</u>
1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature <u>[Signature]</u>	
(OVER)	

*AMB*  
*2/22/10*

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program



Subscribed and sworn before me this 23rd day of February, 2010.

[Signature] (Notary Seal Stamp)

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130112 ✓

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130112 ✓

FEBRUARY 09, 2010

OPERATOR'S NAME:  
SMITH JR, BRUCE W  
OPERATOR'S NUMBER: 3522  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI :  
CAL

O.L. #: 0123456  
DEPT/AGENCY: WSG1  
CASE/REPORT: 10020000  
TEST TYPE: U  
ALCO TARGET VALUE: .083 ✓  
ALCO S-NUM 77199

*BWS*  
*2/22/10*

### — BREATH ANALYSIS —

.083 ADJUSTED FOR 30.14 in ✓  
ALCO TARGET .083 18:31  
BLANK TEST .000 18:32  
INTERNAL STANDARD VERIFIED 18:32  
ALCO TV 30.14 in .083 18:32  
BLANK TEST .000 18:33  
SUBJECT SAMPLE *BWS* .000 18:33  
BLANK TEST .000 18:34  
ALCO TV 30.14 in *2/22/10* ✓ .083 18:34  
BLANK TEST .000 18:35

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130112 ✓

FEBRUARY 09, 2010 ✓  
TIME 18:29

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 43c  
BAROMETER: 30.14 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ ¡ ¢