

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130112

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130112

SEPTEMBER 07, 2010

OPERATOR'S NAME:
SMITH JR, BRUCE W
OPERATOR'S NUMBER: 3522
SUBJECT'S LAST NAME:
UOC
SUBJECT'S FIRST NAME/MI :
UOC

O.L. #: 022345
DEPT/AGENCY: WR61
CASE/REPORT: 10090000
TEST TYPE: U
ALCO TARGET VALUE: .083
ALCO S/N: 77199

— BREATH ANALYSIS —

.083 ADJUSTED FOR 30.32 in
ALCO TARGET .004 16:52
BLANK TEST .000 16:53
INTERNAL STANDARD VERIFIED 16:53
ALCO TV 30.32 in .009 16:53
BLANK TEST .000 16:54
SUBJECT SAMPLE .000 16:55
BLANK TEST .000 16:55
ALCO TV 30.32 in .008 16:56
BLANK TEST .000 16:57

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130112

SEPTEMBER 07, 2010
TIME 16:51

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 42c
BAROMETER: 30.32 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~■