VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130113

Supervisor/Operator Performing the Verification Procedure:

Name Kenneth G. Cox ID# 3527 Date 4/26/2012

A

Agency Skagway Police Department Phone # (907) 983-2232

Instrument Location Skagway Police Department 1st and State Street Skagway, AK 99840

B

Alco S/N ~~N4~~ 4N BMB 5/7/12 Target Value .080 High Pressure 800 psi

Alco Test Values .080 (1st Alco) .080 (2nd Alco)

Signature

(OVER)

BMB 5/7/12

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.

(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 3 day of July, ~~2010.~~ 2012 NAK 7/3/12

(Notary Seal Stamp)

~~Daniel L. Nicholas~~ Nikki Roth
NAK 7/3/12
Notary Public, State of Alaska
Commission Expires with Office

NIKKI ROTH
My Commission Expires
NOTARY PUBLIC
STATE OF ALASKA

STATE OF ALASKA
SCIENTIFIC CRIME DETECTION LABORATORY

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130113

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130113

APRIL 26, 2012

BMB 5/7/12

OPERATOR'S NAME:
COX/KENNETH/G
OPERATOR'S NUMBER: 3527
SUBJECT'S LAST NAME:
VERIFIVATION OF CAL
SUBJECT'S FIRST NAME/MI :
L
O.L. #: L
DEPT/AGENCY: SKG1
CASE/REPORT: 2012-0909
TEST TYPE: V
ALCO TARGET VALUE: .080
ALCO S/N: 4N

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 29.68 in

ALCO TARGET	.079	11:01
BLANK TEST	.000	11:02
INTERNAL STANDARD	VERIFIED	11:02
ALCO TV 29.68 in	.080	11:02
BLANK TEST	.000	11:03
SUBJECT SAMPLE	.000	11:04
BLANK TEST	.000	11:05
ALCO TV 29.68 in	.000	11:06
BLANK TEST	.000	11:06

BMB 5/7/12

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130113

APRIL 26, 2012
TIME 10:57

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.68 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~