

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N: <u>130113</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name: <u>Kenneth G. Cox</u>	ID #: <u>3527</u> Date: <u>January 17, 2009</u> ✓
A Agency: <u>Skagway Police Department</u>	Phone #: <u>907 983-2232</u>
Instrument Location <u>Skagway Police Department, 1<sup>st</sup> &amp; State Street Skagway, Alaska 99840</u> ✓	
Alco S/N <u>X173010</u> ✓	Target Value <u>.082</u> ✓ High Pressure <u>450</u>
Alco Test Value Average <u>.081</u> ✓	<u>.081</u> ✓
	1 <sup>st</sup> Alco 2 <sup>nd</sup> Alco ✓
Signature <u>Kenneth G. Cox</u>	✓
(OVER)	

CGB  
1/17/09

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that:

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Kenneth Cox, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

Jeanne Swartz

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

2/3/09  
Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130113 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130113 ✓

JANUARY 17, 2009 ✓

OPERATOR'S NAME: *COB 12/10/09*  
COX, KENNETH G  
OPERATOR'S NUMBER: 3527  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI :  
L  
O.L. #: L  
DEPT/AGENCY: SKG1  
CASE/REPORT: 2009-0077  
TEST TYPE: U ✓  
ALCO TARGET VALUE: .082 ✓  
ALCO S/N: X173010

### — BREATH ANALYSIS —

.082 ADJUSTED FOR 30.24 in  
ALCO TARGET .082 15:15  
BLANK TEST .000 15:16  
INTERNAL STANDARD VERIFIED 15:16  
ALCO TV 30.25 in .081 ✓ 15:17  
BLANK TEST .000 15:18  
SUBJECT SAMPLE *COB 12/10/09* .000 15:18  
BLANK TEST .000 15:19  
ALCO TV 30.24 in .081 ✓ 15:20  
BLANK TEST .000 15:21

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130113

JANUARY 17, 2009 ✓  
TIME 15:10

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 30.24 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!\*"#&?'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJKLMNPOQRSTUVWXYZ[\]^\_`abcdefgijklmno  
pqrstuvwxyz{|}~■