

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 22 2009

DataMaster cdm S/N: 130113

Supervisor/Operator Performing the Verification Procedure:

Name: Kenneth G. Cox ID #: 3527 Date: July 19, 2009 ✓

A Agency: Skagway Police Department Phone #: 907 983-2232

Instrument Location Skagway Police Department, 1st & State Street Skagway, Alaska 99840 ✓

Alco S/N X173010 ✓ Target Value .082 ✓ High Pressure 400

Alco Test Value Average .081 ✓ .082 ✓
1st Alco 2nd Alco

Signature Kenneth G. Cox ✓ CSB
7/22/09

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 24th day of July 2009.

Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130113

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130113

JULY 19, 2009

OPERATOR'S NAME: *Cox/Kenneth/G*
COX/KENNETH/G
OPERATOR'S NUMBER: 3527
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
L
O.L. #: L
DEPT/AGENCY: SKG1
CASE/REPORT: 2009-2635
TEST TYPE: U
ALCO TARGET VALUE: .002 ✓
ALCO S/N: X173010

--- BREATH ANALYSIS ---

.002 ADJUSTED FOR 30.29 in
ALCO TARGET .003 17:06
BLANK TEST .000 17:07
INTERNAL STANDARD VERIFIED 17:07
ALCO TV 30.29 in .001 ✓ 17:08
BLANK TEST .000 17:09
SUBJECT SAMPLE *Cox/Kenneth/G* .000 17:09
BLANK TEST .000 17:10
ALCO TV 30.30 in .002 ✓ 17:11
BLANK TEST .000 17:12

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130113

JULY 19, 2009 ✓
TIME 17:01

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY ✓
PROGRAM: OKAY ✓
SOFTWARE DATE: 02/20/01 ✓
HEATERS
SAMPLE CHAMBER: 49c ✓
BREATH TUBE: 41c ✓
BAROMETER: 30.30 in ✓
FLOW DETECTOR: OKAY ✓
PUMP
HIGH SPEED: OKAY ✓
DETECTOR: OKAY ✓
FILTERS: OKAY ✓
QUARTZ STANDARD: OKAY ✓
CALIBRATION: OKAY ✓

PRINTER TEST ✓

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLmnopqrstuvwxyz{|}~'abcde fghi jklmno
pqrstuvwxyz{|}~