VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N: 130113

Supervisor/Operator Performing the Verification Procedure:

Name: Kenneth G. Cox ID #: 3527 Date: July 20, 2008

A

Agency: Skagway Police Department Phone #: 907 983-2232

Instrument Location Skagway Police Department, 1st & State Street Skagway, Alaska 99840

Alco S/N X173010 Target Value .082 High Pressure 850

Alco Test Value Average .079 (1st Alco) .080 (2nd Alco)

Signature Kenneth G. Cox KGC 7/25/08

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Kenneth Cox , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

7/28/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BTP 03/00

JUL 24 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130113

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130113

JULY 20, 2008 COB 7/25/08

OPERATOR'S NAME:
COX, KENNETH G
OPERATOR'S NUMBER: 3527
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
L
O.L. #: L
DEPT/AGENCY: SKG1
CASE/REPORT: 2008-2329
TEST TYPE: V
ALCO TARGET VALUE: .082
ALCO S/N: X173010

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 29.88 in

ALCO TARGET	.081	02:38
BLANK TEST	.000	02:39
INTERNAL STANDARD	VERIFIED	02:39
ALCO TV 29.88 in	.079	02:40
BLANK TEST	.000	02:41
SUBJECT SAMPLE	.000	02:42
BLANK TEST	.000	02:43
ALCO TV 29.88 in	.080	02:43
BLANK TEST	.000	02:44

COB 7/25/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130113

JULY 20, 2008
TIME 02:46

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.88 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~