

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N: <u>130113</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name: <u>Kenneth G. Cox</u>	ID #: <u>3527</u> Date: <u>September 18, 2008</u> ✓
A Agency: <u>Skagway Police Department</u>	Phone #: <u>907 983-2232</u>
Instrument Location <u>Skagway Police Department, 1<sup>st</sup> &amp; State Street Skagway, Alaska 99840</u> ✓	
Alco S/N <u>X173010</u> ✓	Target Value <u>.082</u> ✓ High Pressure <u>600 psi</u>
Alco Test Value Average <u>.079</u> ✓	<u>.080</u> ✓
1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature <u>Kenneth G. Cox</u> ✓	<u>CSB</u> <u>9/18/08</u>
(OVER)	


(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Kenneth Cox, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

10/15/08

Date

SEP 22 2008

BT9 03/00

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130113 *e*

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130113

SEPTEMBER 18, 2008 *CSB 9/20/08*

OPERATOR'S NAME:  
COX, KENNETH G  
OPERATOR'S NUMBER: 3527  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI :

L  
O.L. #: L  
DEPT/AGENCY: SKG1  
CASE/REPORT: 2008-3687  
TEST TYPE: U  
ALCO TARGET VALUE: .082 ✓  
ALCO S/N: X173010 ✓

### --- BREATH ANALYSIS ---

.082 ADJUSTED FOR 30.04 in		
ALCO TARGET	.082	00:45
BLANK TEST	.000	00:46
INTERNAL STANDARD	VERIFIED	00:46
ALCO TV 30.04 in	.079 ✓	00:47
BLANK TEST	.000	00:48
SUBJECT SAMPLE	.000 <i>CSB 9/20/08</i>	00:48
BLANK TEST	.000	00:49
ALCO TV 30.04 in	.080 ✓	00:50
BLANK TEST	.000	00:50

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130113

SEPTEMBER 18, 2008  
TIME 00:53

### --- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	40c
BAROMETER:	30.04 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHI  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopq  
rstuvwxyz{|}~