

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

OCT 26 2009

DataMaster cdm S/N: <u>130113</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name: <u>Kenneth G. Cox</u>	ID #: <u>3527</u> Date: <u>November 1, 2009</u> ✓
A Agency: <u>Skagway Police Department</u>	Phone #: <u>907 983-2232</u>
Instrument Location <u>Skagway Police Department, 1st & State Street Skagway, Alaska 99840</u> ✓	
Alco S/N <u>X173010</u>	Target Value <u>.082</u> High Pressure <u>300</u>
Alco Test Value Average <u>.081</u>	<u>.082</u>
1 st Alco	2 nd Alco
Signature <u><i>Kenneth G. Cox</i></u>	
(OVER)	

(Do Not write in the area below)

BMB
11/1/09

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 21st day of Dec, 2009.

Carolyn M. Noland
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130113

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130113

NOVEMBER 01, 2009

OPERATOR'S NAME:
COX/KENNETH/S

OPERATOR'S NUMBER: 3527

SUBJECT'S LAST NAME:

VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :

L
O.L. #: L

DEPT/AGENCY: SKG1

CASE/REPORT: 2009-4315

TEST TYPE: U

ALCO TARGET VALUE: .002 ✓

ALCO S/N: X173010 ✓

--- BREATH ANALYSIS ---

.002 ADJUSTED FOR 30.04 in
ALCO TARGET .002 16:55
BLANK TEST .000 16:56
INTERNAL STANDARD VERIFIED 16:56
ALCO TV 30.04 in .001 16:57
BLANK TEST .000 16:58
SUBJECT SAMPLE .000 16:58
BLANK TEST .000 16:59
ALCO TV 30.04 in .002 17:00
BLANK TEST .000 17:00

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130113

NOVEMBER 01, 2009
TIME 17:02

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHOUSER: 49c
BREATH TUBE: 42c
BAROMETER: 30.04 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./:;<=>?@ABCDEFGHIJKL
HIJKLMNOPQRSTUVWXYZ0123456789:;
=abcdefghijklmnopqrstuvwxyz{|}~