

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130113 ✓
Name		Kenneth G. Cox	ID# 3527 ✓ Date 10/21/2010 ✓
A	Agency	Skagway Police Department	Phone # (907) 983-2232
Instrument Location Skagway Police Department 1st and State Street Skagway, AK 99840			
Alco S/N		X173019 ✓	Target Value .082 ✓ High Pressure 900
Alco Test Values		.079 ✓ 1st Alco	.080 ✓ 2nd Alco
Signature		<i>Kenneth G. Cox</i>	

(OVER)

*BMB
12/9/10*

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 21ST day of DEC, 2010.

Daniel L. Nicholas (Notary Seal Stamp)

Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130113

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130113

OCTOBER 21, 2010

OPERATOR'S NAME: COX/KENNETH/G
OPERATOR'S NUMBER: 3527
SUBJECT'S LAST NAME: VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI : L
O.L. #: L
DEPT/AGENCY: SK61
CASE REPORT: 2010
TEST TYPE: V
ALCO TARGET VALUE: .002
ALCO S/N: X173019

— BREATH ANALYSIS —

.002 ADJUSTED FOR 30.04 in
ALCO TARGET .002 16:27
BLANK TEST .000 16:28
INTERNAL STANDARD VERIFIED 16:28
ALCO TV 30.04 in .079 16:28
BLANK TEST .000 16:29
SUBJECT SAMPLE .000 16:30
BLANK TEST .000 16:30
ALCO TV 30.04 in .000 16:31
BLANK TEST .000 16:32

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130113

OCTOBER 21, 2010
TIME 16:23

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.04 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abdefghijklmno
pqrstuvwxyz{|}~■