

APR 04 2011

VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument
State of Alaska*

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>15044 130114</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Kevin Stevenson</u>	ID# <u>2414</u> Date <u>04.1.11</u>
A	
Agency <u>WAA Police Dept</u>	Phone # <u>907-786-1120</u>
Instrument Location <u>WAA Police Dept Eugene Short Hall 2601 Providence Dr</u>	
B	
Alco S/N <u>630491</u>	Target Value <u>.079</u> High Pressure <u>1000</u>
Alco Test Values	<u>.079</u> <u>.078</u>
1st Alco	2nd Alco
Signature <u><i>[Handwritten Signature]</i></u>	

(OVER)

(Do Not write in the area below)

[Handwritten initials]
4/4/11

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 31 day of MAY, 2010. 2011



[Handwritten Signature] (Notary Seal Stamp)
Daniel L. Nicholas Kyle Klopfer
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130114

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130114

APRIL 01, 2011

OPERATOR'S NAME:
STEVENS ON KEVIN L
OPERATOR'S NUMBER: 2414
SUBJECT'S LAST NAME:
VERIFICATION
SUBJECT'S FIRST NAME/NI :
VERIFICATION
O.L. #: 231232
DEPT/AGENCY: ANCF
CASE/REPORT: 12321
TEST TYPE: U
ALCO TARGET VALUE: .079
ALCO S/N: 630491

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.85 in
ALCO TARGET .078 06:12
BLANK TEST .000 06:13
INTERNAL STANDARD VERIFIED 06:13
ALCO TV 29.85 in .079 06:14
BLANK TEST .000 06:14
SUBJECT SAMPLE .000 06:15
BLANK TEST .000 06:16
ALCO TV 29.85 in .078 06:16
BLANK TEST .000 06:17

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130114

APRIL 01, 2011
TIME 06:10

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 29.85 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~@

