

VERIFICATION OF CALIBRATION REPORT

MAY 04 2009

of DataMaster edm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster edm S/N 130114 ^{COB 5/13/09}

Name Kevin Stevenson ID# 2414 Date 04-26-09

A Agency WAA Police Dept Phone # 907-786-1120

Instrument Location WAA Police Dept ✓

B Alco S/N X 301591 ✓ Target Value .080 ✓ High Pressure 300

Alco Test Value Average .081 ✓
1st Alco .081 ✓
2nd Alco

Signature [Handwritten Signature] ✓
COB
5/18/09

(OVER)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Handwritten Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22 day of May, 2009.

[Handwritten Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



MAY 04 2009

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130114 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130114

APRIL 26, 2009 ✓ *COB strip*

OPERATOR'S NAME:
STEVENSON KEVIN L
OPERATOR'S NUMBER: 2414
SUBJECT'S LAST NAME:
VERIFICATION
SUBJECT'S FIRST NAME/MI :
OO L
O.L. #: 1234
DEPT/AGENCY: RMC7
CASE/REPORT: 0900000
TEST TYPE: U
ALCO TARGET VALUE: .000 ✓

BREATH ANALYSIS

.000 ADJUSTED FOR 30.58 in ✓
ALCO TARGET .001 23:15
BLANK TEST .000 23:16
INTERNAL STANDARD VERIFIED 23:16
ALCO TV 30.58 in .001 ✓ 23:16
BLANK TEST .000 23:17
SUBJECT SAMPLE *COB strip* .000 23:18
BLANK TEST .000 23:19
ALCO TV 30.58 in .001 ✓ 23:19
BLANK TEST .000 23:20

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130114

APRIL 26, 2009 ✓
TIME 23:12

DIAGNOSTIC CHECK

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/28/01
METERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.58 in

PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

*002*0**..0123456789:*(=)90000000
HELLO\WORLD\STUART\KYLE J. *state\tyhi.johann
pqr\lmnopqz(!)-@