

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JAN 13 2010


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130114</u>
Name <u>Kevin Stevenson</u>	ID# <u>2414</u>	Date <u>01-08-10</u> <u>01-07-10</u> (1-12-10 CWJ)
A Agency <u>UAA Police Dept</u>	Phone # <u>907-786-1120</u>	
Instrument Location <u>3211 Providence Dr. suite 114 ESH, Anchorage, AK</u>		
Alco S/N <u>75470</u>		Target Value <u>.080</u> High Pressure <u>700</u>
Alco Test Value Average <u>.080</u>		<u>.081</u>
1 <sup>st</sup> Alco		2 <sup>nd</sup> Alco
Signature <u>[Signature]</u>		<u>CMB</u> <u>1/13/10</u>
(OVER)		

(Do Not write in the area below)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
  - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
  - (3) I am the Scientific Director of the State Breath Alcohol Program.
  - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
  - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
  - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 26<sup>th</sup> day of January, 2010.  
[Signature] (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JAN 12 2010

DataMaster cdm S/N 130114 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130114 ✓

JANUARY 08, 2010 ✓ *CSB*

OPERATOR'S NAME:  
STEINBSON KEVIN L  
OPERATOR'S NUMBER: 2414  
SUBJECT'S LAST NAME:  
VERIFICATION  
SUBJECT'S FIRST NAME/MI :  
VERIFICATION  
O.L. #: 12345  
DEPT/AGENCY: 0007  
CASE/REPORT: 10001  
TEST TYPE: U  
ALCO TARGET VALUE: .000 ✓  
ALCO S/N: 75470 ✓

### --- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.55 in ✓  
ALCO TARGET .079 00:29  
BLANK TEST .000 00:29  
INTERNAL STANDARD VERIFIED 00:29  
ALCO TU 29.55 in .000 ✓ 00:29  
BLANK TEST .000 00:30  
SUBJECT SAMPLE *CSB* .000 00:31  
BLANK TEST .000 00:32  
ALCO TU 29.55 in .001 ✓ 00:32  
BLANK TEST .000 00:33

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130114 ✓

JANUARY 08, 2010 ✓  
TIME 00:25 ✓

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 29.55 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@BCDEFG  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~!@#