

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130114

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130114

JULY 08, 2010

OPERATOR'S NAME:

STEVENSON KEVIN L

OPERATOR'S NUMBER: 2414

SUBJECT'S LAST NAME:

VERIFICATION

SUBJECT'S FIRST NAME/MI:

VERIFICATION

D.L. #: 1034

DEPT/AGENCY: ANCT

CASE/REPORT: 01098

TEST TYPE: U

ALCO TARGET VALUE: .080

ALCO S/N: 75470

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 30.11 in
ALCO TARGET .080 22:14
BLANK TEST .000 22:15
INTERNAL STANDARD VERIFIED 22:15
ALCO TO 30.11 in .082 22:16
BLANK TEST .000 22:17
SUBJECT SAMPLE .000 22:17
BLANK TEST .000 22:18
ALCO TO 30.11 in .082 22:19
BLANK TEST .000 22:19

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130114

JULY 08, 2010

TIME 22:12

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/08/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 30.11 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNORSTUVWXYZ[\]^_`abcde fghijklmno
pqrs tuvwxyz{|}~