

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: **DataMaster cdm S/N** 130114 ✓

Name Kevin Stevenson ID# 2414 ✓ Date 9/9/10 ~~9/13/10~~

A Agency WAAPD Phone # 786-1120

Instrument Location WAAPD Suite 114 ESH

B Alco S/N 630491 ✓ Target Value .079 ✓ High Pressure 1400

Alco Test Values .078 ✓ 1st Alco .078 ✓ 2nd Alco

Signature Kevin Stevenson

(OVER)

Handwritten initials and date: 9/17/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Signature of Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program



Subscribed and sworn before me this 9th day of NOV, 2010.

Signature of Daniel L. Nicholas (Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130114

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DataMaster cdm 130114

SEPTEMBER 09, 2010

OPERATOR'S NAME: STEVENSON KEVIN L
OPERATOR'S NUMBER: 1414
SUBJECT'S LAST NAME: VERIFICATION
SUBJECT'S FIRST NAME:

ALCOHOL TARGET VALUE: .079
ALCOHOL S/N: 630491

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.81 in	
ALCOHOL TARGET	.079
BLANK TEST	000
INTERNAL STANDARD	000
ALCOHOL TARGET	.079
BLANK TEST	000
INTERNAL STANDARD	000
ALCOHOL TARGET	.079
BLANK TEST	000
INTERNAL STANDARD	000

BT9 06/03

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DataMaster cdm 130114

SEPTEMBER 09, 2010
TIME 20:16

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/09/08

HEATERS: OKAY
SAMPLE CHAMBER: OKAY
BREATH TUBE: OKAY

BAROMETRIC: 29.31 in

FLOW DETECTOR: OKAY

PUMP: OKAY
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

BT9 06/03