

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument*

*State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>13011-1</u>	
A	Name <u>Kevin Stevenson</u>	ID# <u>2414</u>	Date <u>12-31-10</u>
	Agency <u>UAA Police Dept</u>	Phone # <u>907-786-1120</u>	
	Instrument Location <u>UAA Police Dept</u>		
B	Alco S/N <u>630491</u>	Target Value <u>± 0.09</u>	High Pressure <u>1100</u>
	Alco Test Values	<u>0.78</u>	<u>.078</u>
	1 <sup>st</sup> Alco		2 <sup>nd</sup> Alco
	Signature <u>[Handwritten Signature]</u>		<u>COB</u> <u>1/18/11</u>

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


[Handwritten Signature]  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 9<sup>th</sup> day of MAR, 2011

[Handwritten Signature] (Notary Seal Stamp)

Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office

BT9 06/03





IR1 0.094 IR2 0.093 COMMENT (IR - TV @ BP) = -.008

# VERIFICATION OF CALIBRATION REPORT

## of DataMaster cdm Breath Test Instrument

### State of Alaska

#### Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

FC1R1DIFF FC2R1DIFF TESTTIME 6

DataMaster cdm S/N 130114 ✓

ISN: 130180 LOCATION: Bristol Bay AWT INPUTDATE 11/27/2001 7:53:19 AM RECORD 6011

OPERATOR 2467 Name Kevin Stevenson ID# 2414 ✓ Date 12-31-10 ✓  
CERT DATE 11/27/2001 Date to be signed: DIRECTOR

ALCOSN: X124662 Agency WAA Police Dept Phone # 907-786-1120  
TARGET: 0.102 PRESSURE: 700

Chan 0: Instrument Location WAA Police Dept

Chan 1: Alco S/N 630491 Target Value .079 ✓ High Pressure 1100

Chan 2: Delete and Exit Print VOC Save and Exit  
Chan 3: Alco Test Values .078 ✓ .078 ✓  
1st Alco 2nd Alco

IR1 0.092 IR2 0.096 COMMENT (IR - TV @ BP) = -.004

Signature: *[Handwritten Signature]*

(OVER)

FC1R1DIFF FC2R1DIFF (Do Not Write in the area below)

ISN: 130180 LOCATION: SELDOVIA PD INPUTDATE 2/7/2002 11:06:40 AM RECORD 6111

OPERATOR 2467 CERT DATE 1/14/2002 Date to be signed: DIRECTOR

ALCOSN: X124662 TARGET: 0.102 PRESSURE: 700

Chan 0:  
Chan 1:  
Chan 2: Delete and Exit Print VOC Save and Exit  
Chan 3:

IR1 0.096 IR2 0.096 COMMENT (IR - TV @ BP) = -.007

FC1 FC2

FC1PEAK FC2PEAK

BASELN1 BASELN2

FC1R1DIFF FC2R1DIFF TESTTIME 6

ISN: 130180 LOCATION: SELDOVIA PD INPUTDATE 3/14/2002 4:01:38 PM RECORD 6165

OPERATOR 2467 CERT DATE 3/6/2002 Date to be signed: DIRECTOR

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N ~~130114~~ 130114

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130114

DECEMBER 31, 2010

OPERATOR'S NAME: STEVENSON KEVIN L  
OPERATOR'S NUMBER: 2414  
SUBJECT'S LAST NAME: VERIFICATION  
SUBJECT'S FIRST NAME/MI: VERIFICATION  
O.L. #: 12321  
DEPT/AGENCY: ANC7  
CASE/REPORT: 10302  
TEST TYPE: U  
ALCO TARGET VALUE: .079  
ALCO S/N: 630491

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.64 in  
ALCO TARGET .078 22:54  
BLANK TEST .000 22:55  
INTERNAL STANDARD VERIFIED 22:55  
ALCO TV 29.64 in .078 22:55  
BLANK TEST .000 22:56  
SUBJECT SAMPLE .000 22:57  
BLANK TEST .000 22:58  
ALCO TV 29.64 in .078 22:58  
BLANK TEST .000 22:59

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130114

DECEMBER 31, 2010  
TIME 22:51

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 29.64 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF  
HIJKLNPQRSTUWXYZ[\]^\_`abcde fghi.jk lmnop  
pqrstuvwxyz{|}~■