

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAY 18 2011

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130115

Name DOUG E. HARRIS ID# 5530 Date MAY 15, 2011

A Agency BRESTOL Bay Borough PD Phone # 246-4222

Instrument Location BLDA# 150 KSAB King Salmon AK

B Alco S/N 31085 Target Value 075 High Pressure _____

Alco Test Values 072 072
1st Alco 2nd Alco

Signature Doug E. Harris

(OVER)

BMB
5/18/11

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 15 day of JUNE, 2011. NJB

Daniel E. Nicholas (Notary Seal Stamp)
Daniel E. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130115

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130115

MAY 15, 2011

OPERATOR'S NAME:
HARRIS DOUG E
OPERATOR'S NUMBER: 5530
SUBJECT'S LAST NAME:
TEST
SUBJECT'S FIRST NAME/MI :
TEST

D.L. #: 01010101
DEPT/AGENCY: NKN1
CASE/REPORT: 0010010
TEST TYPE: U
ALCO TARGET VALUE: 075
ALCO S/N: 31085

--- BREATH ANALYSIS ---

.075 ADJUSTED FOR 29.44 in
ALCO TARGET .073 16:07
BLANK TEST .000 16:08
INTERNAL STANDARD VERIFIED 16:08
ALCO TO 29.44 in .072 16:09
BLANK TEST .000 16:10
SUBJECT SAMPLE .000 16:10
BLANK TEST .000 16:11
ALCO TO 29.44 in .072 16:11
BLANK TEST .000 16:12

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130115

MAY 15, 2011
TIME 16:12

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 29.44 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

"000"()!+,-./0123456789:;{=}?@ABCDEF6
HIJKL MNOPQRSTUVWXYZ * abcdefghijklmno
pqrstuvwxyz{|}~■