

MAR 31 2009

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 13045

Name Jerry Castleberry ID# 2036 Date 3/26/09

A Agency Bristol Bay Borough Police Phone # 246-4222

Instrument Location King Salmon

B Alco S/N X172961 Target Value .082 High Pressure 500

Alco Test Value Average 0.080 1.80 ms 0.080 1.80 ms

1st Alco 2nd Alco

Signature Jerry E Castleberry CDS
4/14/09

Under the Alaska Rules of Evidence, I certify that (OVER)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Jerry Castleberry, whose name appears in my records, is qualified to verify the calibration of the above instrument.


(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

4/14/09
Date



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130115

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130115

MARCH 26, 2009

OPERATOR'S NAME: **COB**
CASTLEBERRY JERRY
OPERATOR'S NUMBER: 0035
SUBJECT'S LAST NAME:
VERIFICATION CAL
SUBJECT'S FIRST NAME: N/A
G.L. #: 3009
TEST AGENCY: NKV
CASE/REPORT: 09227
TEST TYPE: V
BLOOD TARGET VALUE: 0.02
ALCO CAL: N17291

--- BREATH ANALYSIS ---

ALCO ADJUSTED FOR 29.20 in
ALCO TEST .000 08:56
BLANK TEST .000 08:56
INTERNAL STANDARD VERIFIED 08:57
ALCO TO 29.20 in .000 08:57
BLANK TEST .000 08:58
SUBJECT SAMPLE **COB** .000 08:58
BLANK TEST .000 08:59
ALCO TO 29.20 in .000 08:59
BLANK TEST .000 09:00

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130115

MARCH 26, 2009
TIME 20:52

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 45c
BREATH TUBE: 41c
BAROMETER: 29.20 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

ALCO (1) = 0.000, ALCO (2) = 0.000, ALCO (3) = 0.000
ALCO (4) = 0.000, ALCO (5) = 0.000, ALCO (6) = 0.000
ALCO (7) = 0.000, ALCO (8) = 0.000, ALCO (9) = 0.000