

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130115
Name Kevin Oldham ID# 5227 Date 5/9/08
A Agency Bristol Bay Borough PD Phone # 246-4222
Instrument Location King Salmon

B Alco S/N X172961 Target Value .082 High Pressure 120
Alco Test Value Average .081 .081
1st Alco 2nd Alco

Signature Kevin Oldham COB
6/5/08

Under the Alaska Rules of Evidence, I certify that;


(OVER)

(a) I am Jeanne Swartz, Forensic Scientist III, ~~Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.~~
(Do Not Write in the area below)

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Kevin Oldham, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz 6/5/08
Date
Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

MAY 14 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130115

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130115

MAY 09, 2008

OPERATOR'S NAME: *CAS 6/3/08*
OLDHAM, JAMES KEVIN
OPERATOR'S NUMBER: 5227
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L.#: VOC
DEPT/AGENCY: NK41
CASE/REPORT: VOC
TEST TYPE: V
ALCO TARGET VALUE: .002
ALCO S/N: K172631

--- BREATH ANALYSIS ---

80/21

.002 ADJUSTED FOR 29.75 in		
ALCO TARGET	.001	21:00
BLANK TEST	.000	21:01
INTERNAL STANDARD	VERIFIED	21:01
ALCO TO 29.75 in	.001	21:02
BLANK TEST	.000	21:02
SUBJECT SAMPLE	.000	21:03
BLANK TEST	.000	21:04
ALCO TO 29.77 in	.001	21:04
BLANK TEST	.000	21:05

CAS 6/3/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130115

MAY 09, 2008
TIME 20:57

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 45c
BREATH TUBE: 49c
BAROMETER: 29.75 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKL MNOPQRSTU VWXYZ[\]^_`ab cdefghijklmno
pqrstuvwxyz{|}~