

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 07 2009

130115
6-16-09
CMB

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N X172961

Name DOUG HARRIS ID# 5530 Date JUNE 5, 2009

A Agency BRISTOL BAY BOROUGH POLICE NKN1 Phone # 907-246-4222

Instrument Location KENA SALMON, AK ✓

B Alco S/N X172961 ✓ Target Value .082 ✓ High Pressure 100

Alco Test Values .082 ✓ .082 ✓
1st Alco 2nd Alco

Signature Doug E. Harris

(OVER)

BMB
6/16/09

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 24th day of June, 2009.
Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

130115 BMB
6/16/09

DataMaster cdm S/N X172961

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
SCIENTIFIC CRIME DETECTION LABORATORY
1301 W. BRIDGEWAY
ANCHORAGE, ALASKA 99501
PHONE: (907) 261-2000
FAX: (907) 261-2001
WWW.AKSTATEPOLICE.COM

TEST NO. 130115
SUBJECT NAME
TEST DATE 6/16/09
TEST TIME
TESTER
INSTRUMENT
CALIBRATION DATE
CALIBRATION VALUE
CALIBRATION S/N

TEST ANALYSIS

130115	✓
130116	
130117	
130118	
130119	
130120	
130121	
130122	
130123	
130124	
130125	
130126	
130127	
130128	
130129	
130130	

TESTER SIGNATURE
DATE

✓ BMB
6/16/09

✓ BMB
6/16/09

ALASKA DEPARTMENT OF PUBLIC SAFETY
SCIENTIFIC CRIME DETECTION LABORATORY
1301 W. BRIDGEWAY
ANCHORAGE, ALASKA 99501
PHONE: (907) 261-2000
FAX: (907) 261-2001
WWW.AKSTATEPOLICE.COM

TEST NO. 130115
SUBJECT NAME
TEST DATE 6/16/09
TEST TIME
TESTER
INSTRUMENT
CALIBRATION DATE
CALIBRATION VALUE
CALIBRATION S/N

TEST ANALYSIS

130115	✓
130116	
130117	
130118	
130119	
130120	
130121	
130122	
130123	
130124	
130125	
130126	
130127	
130128	
130129	
130130	

TESTER SIGNATURE
DATE