

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

SEP 25 2008

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N 130115 130115 ✓
A	Name <u>Jerry Castleberry</u>	ID# <u>2636</u> Date <u>9/19/08</u> ✓
	Agency <u>Bristol Bay Borough Police</u>	Phone # <u>246-4222</u>
	Instrument Location <u>King Salmon</u> ✓	
B	Alco S/N <u>X172941</u> ✓	Target Value <u>.082</u> ✓ High Pressure <u>800</u>
	Alco Test Value Average <u>.081</u>	<u>.081, .080</u>
	1 st Alco	2 nd Alco <u>CMW 9-26-08</u>
	Signature <u>Jerry Castleberry</u> ✓	<u>CSB</u> <u>9/30/08</u>


Under the Alaska Rules of Evidence, I certify that, (OVER)

(Do Not write in the area below)
(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Jerry Castleberry, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

10/5/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

80712 2243

DataMaster cdm S/N 130115 ✓

(CONTINUED FROM FRONT PAGE)

C ~~TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.~~

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130115

SEPTEMBER 19, 2008

OPERATOR'S NAME:
CASTLEBERRY JERRY
OPERATOR'S NUMBER: 0635
SUBJECT'S LAST NAME:
VERIFICATION CAL
SUBJECT'S FIRST NAME/MI :
LO

O.L. #: 909
DEPT/AGENCY: NKA1
CASE/REPORT: 00000
TEST TYPE: U
ALCO TARGET VALUE: .082 ✓
ALCO S/N: X172951 ✓

--- BREATH ANALYSIS ---

ALCO ADJUSTED FOR 29.65 in 13:41
ALCO TAPSET - .081 13:42
BLANK TEST .000 13:42
INTERNAL STANDARD VERIFIED .081 13:42
ALCO TO 29.65 in .081 13:43
BLANK TEST .000 13:44
SUBJECT SAMPLE .300 13:44
BLANK TEST .000 13:45
ALCO TO 29.65 in .080 13:45
BLANK TEST .000 13:46

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130115

SEPTEMBER 19, 2008 ✓
TIME 13:37

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 49c
BAROMETER: 29.65 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKL MNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~