VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

OCT 26 2009

11-09 CMJ

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130115

A

Name DOUG HARRIS ID# 5530 Date OCT 16, 09

Agency Bristol Bay Police Phone # 246-4222

Instrument Location Bristol Bay Borough Police Dept

B

Alco S/N X124623 Target Value .076 High Pressure

Alco Test Values .075 (1st Alco) .075 (2nd Alco)

Signature Doug Harris

(OVER)

BMB 11/17/09

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 3rd day of Dec, 2009.

(Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

Carolyn M. Noland NOTARY PUBLIC State of Alaska

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

130115 BMB

DataMaster cdm S/N ~~X124623~~ 11/14/09

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130115

OCTOBER 16, 2009 BMB 11/14/09

OPERATOR'S NAME:
MORRIS DOUG E
OPERATOR'S NUMBER: 3530
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: 000000
DEPT/AGENCY: NKN1
CASE/REPORT: TEST
TEST TYPE: V
ALCO TARGET VALUE: .076
ALCO S/N: X124623

— BREATH ANALYSIS —

.076 ADJUSTED FOR 29.60 in

ALCO TARGET	.075	20:39
BLANK TEST	.000	20:40
INTERNAL STANDARD	VERIFIED	20:41
ALCO TU 29.60 in	.075	20:41
BLANK TEST	.000	20:42
SUBJECT SAMPLE BMB 11/14/09	.000	20:42
BLANK TEST	.000	20:43
ALCO TU 29.60 in	.075	20:43
BLANK TEST	.000	20:44

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130115

OCTOBER 16, 2009

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.60 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST