

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>13045</u> ✓
A	Name	<u>Jerry Castleberry</u>	ID# <u>2636</u> Date <u>11/25/08</u> ✓
	Agency	<u>Bristol Bay Borough Police</u>	Phone # <u>246-4222</u>
	Instrument Location	<u>King Salmon</u> ✓	
B	Alco S/N	<u>X172961</u> ✓	Target Value <u>.082</u> ✓ High Pressure <u>700</u> ✓
	Alco Test Value Average	<u>.081</u> ✓ 1 <sup>st</sup> Alco	<u>.081</u> ✓ 2 <sup>nd</sup> Alco
Signature		<u>Jerry Castleberry</u> ✓	<u>CSB</u> <u>12/1/08</u>

Under the Alaska Rules of Evidence, I certify that (OVER)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Jerry Castleberry, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

Jeanne Swartz 12/4/08  
Date  
Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

DEC 01 2008

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130115

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130115

NOVEMBER 25, 2008

OPERATOR'S NAME: CASTLEBERRY JERRY E  
OPERATOR'S NUMBER: 2636  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI:  
CAL

O.L. #: 09090  
DEPT/AGENCY: NKNM  
PAGE/REPORT: 09/8  
TEST TYPE: U  
ALCO TARGET VALUE: .082 ✓  
ALCO S/N: X172961

### --- BREATH ANALYSIS ---

.082 ADJUSTED FOR 29.52 in  
ALCO TARGET .080 16:07  
BLANK TEST .000 16:08  
INTERNAL STANDARD VERIFIED 16:08  
ALCO TU 29.52 in .081 ✓ 16:09  
BLANK TEST .000 16:09  
SUBJECT SAMPLE .082 16:10  
BLANK TEST .000 16:11  
ALCO TU 29.54 in .081 ✓ 16:11  
BLANK TEST .000 16:12

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130115

NOVEMBER 25, 2008  
TIME 16:13

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 95c  
BREATH TUBE: 91c  
BAROMETER: 29.54 in  
FLOW DETECTOR: OKAY  
PUMP HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

ALCO TARGET: .082  
ALCO TU: .081  
ALCO S/N: X172961