


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

AUG 16 2010


Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

| | | | |
|--|-------------|---|----------------------------------|
| Supervisor/Operator Performing the Verification Procedure: | | DataMaster cdm S/N | 130115 ^{OWS} 8/16/10 |
| Name | Doug Harris | ID# | 5530 ✓ Date Aug 8, 2010 ✓ |
| Agency | BBPD | Phone # | 246-4222 |
| Instrument Location King SALMON AK BBPD | | | |
| Alco S/N | | X301605 ✓ | Target Value 075 ✓ High Pressure |
| Alco Test Values | | 074 ✓ 1 st Alco | 073 ✓ 2 nd Alco |
| Signature | |  COB 8/7/10 | |
| (OVER) | | | |


(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 23rd day of SEPT, 2010.


(Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130115

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130115

AUGUST 08, 2010

OPERATOR'S NAME:

HARRIS DOUG E

OPERATOR'S NUMBER: 5539

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI :

TEST

D.L. #: 0010101

DEPT/AGENCY: NKN1

CASE/REPORT: XXXXXX

TEST TYPE: U

ALCO TARGET VALUE: 075

ALCO S/N: X301635

--- BREATH ANALYSIS ---

.075 ADJUSTED FOR 29.47 in
ALCO TARGET .073 15:27
BLANK TEST .000 15:28
INTERNAL STANDARD VERIFIED 15:28
ALCO TO 29.47 in .074 15:29
BLANK TEST .000 15:29
SUBJECT SAMPLE .000 15:30
BLANK TEST .000 15:31
ALCO TO 29.47 in .073 15:31
BLANK TEST .000 15:32

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130115

AUGUST 08, 2010

TIME 15:33

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 49c

BAROMETER: 29.47 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKL MNOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~||