

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 13 2011

DataMaster cdm S/N 130116

Supervisor/Operator Performing the Verification Procedure:  
Name STEVE KEVAN ID# 6585 Date 5/30/11

**A** Agency ALASKA STATE PROSPECTORS Phone # 907-438-6500  
Instrument Location EMMONAK ASI

**B** Alco S/N 77111 Target Value .098 High Pressure 650  
Alco Test Values 1077 .097  
1st Alco 2nd Alco

Signature [Signature]

(OVER)

BMB  
6/13/11

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 16 day of JUNE, 2011. NJB

[Signature] (Notary Seal Stamp)  
Daniel L. Nicholas Kyle Klopfer  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130116

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130116

MAY 30, 2011

OPERATOR'S NAME: KEVAN, STEVE  
OPERATOR'S NUMBER: 6585  
SUBJECT'S LAST NAME: TEST  
SUBJECT'S FIRST NAME/NI: TEST  
O.L. #: 00  
DEPT/AGENCY: EMMO  
CASE/REPORT: 11-  
TEST TYPE: U  
ALCO TARGET VALUE: .078  
ALCO S/N: 77111

### --- BREATH ANALYSIS ---

.078 ADJUSTED FOR 30.48 in  
ALCO TARGET .079 10:17  
BLANK TEST .000 10:18  
INTERNAL STANDARD VERIFIED 10:18  
ALCO TV 30.48 in .077 10:18  
BLANK TEST .000 10:19  
SUBJECT SAMPLE .000 10:19  
BLANK TEST .000 10:20  
ALCO TV 30.48 in .077 10:21  
BLANK TEST .000 10:22

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130116

MAY 30, 2011  
TIME 10:13

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 40c  
BAROMETER: 30.48 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@BCDEFG  
HIJKLMNORSTUVWXYZ[\]^\_`abcdefgijklmno  
pqrstuvwxyz{|}~■