FEB 0 4 2009

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130116

A
Name Kenneth Parker ID# 5578 Date 01-08-09
Agency McGrath Public Safety Phone # 524-3025
Instrument Location Capt Snow Building AST office

B
Alco S/N 77111 Target Value .078 High Pressure .750

Alco Test Values .076 (1st Alco) .075 (2nd Alco)

Signature [signature] 2/9/09

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Kenneth Parker , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed

[signature] 2/9/09

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BT9 03/00

FEB 0 5 2009

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130116

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130116

JANUARY 08, 2009

COB 2/9/09

OPERATOR'S NAME:
PARKER/KENNETH/W
OPERATOR'S NUMBER: 5578
SUBJECT'S LAST NAME:
PARKER
SUBJECT'S FIRST NAME/MI :
KENNETH/W
O.L. #: 130116
DEPT/AGENCY: MCG1
CASE/REPORT: T
TEST TYPE: T
ALCO TARGET VALUE: .078
ALCO S/N: 77111

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.70 in

ALCO TARGET	.077	11:05
BLANK TEST	.000	11:06
INTERNAL STANDARD	VERIFIED	11:06
ALCO TV 29.70 in	.076	11:07
BLANK TEST	.000	11:08
SUBJECT SAMPLE	.000	11:08
BLANK TEST	.000	11:09
ALCO TV 29.70 in	.075	11:09
BLANK TEST	.000	11:10

COB 2/9/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130116

JANUARY 08, 2009
TIME 10:58

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	42c
BAROMETER:	29.70 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

```
 !"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~
```