

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130116

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130116

APRIL 10, 2008

OPERATOR'S NAME: PARKER KENNETH W
OPERATOR'S NUMBER: 5578
SUBJECT'S LAST NAME: PARKER

SUBJECT'S FIRST NAME/MI: KENNETH W

O.L. #: 130116
DEPT/AGENCY: A661
CASE/REPORT: T
TEST TYPE: T
ALCO TARGET VALUE: 078
ALCO S/N: 77111

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.83 in		
ALCO TARGET	.077	12:51
BLANK TEST	.000	12:52
INTERNAL STANDARD	VERIFIED	12:52
ALCO TU 29.83 in	.076	12:52
BLANK TEST	.000	12:53
SUBJECT SAMPLE	.000	12:54
BLANK TEST	.000	12:55
ALCO TU 29.83 in	.075	12:55
BLANK TEST	.000	12:56

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130116

APRIL 10, 2008
TIME 12:47

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.83 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghi jk lmn
opq rstuvwxyz{|}~■