

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130116 ✓
Name	Kenneth Parker	ID#	5578 Date 10-24-08
A Agency	McGrath Public Safety	Phone #	524-3095
Instrument Location	McGrath AST ✓		
B Alco S/N		7714 ✓	Target Value 0.78 ✓ High Pressure 600 -
Alco Test Values		0.75 ✓	0.75 -
		1 st Alco	2 nd Alco
Signature			CBS 11/4/08

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Kenneth Parker, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

11/5/08
Date

BT9 03/00

OCT 30 2008

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm breath Test Instrument
 State of Alaska
 Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130116 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130116

OCTOBER 24, 2008 ✓ *CSB 11/4/08*

OPERATOR'S NAME: PARKER/KENNETH/W
 OPERATOR'S NUMBER: 5578
 SUBJECT'S LAST NAME: PARKER
 SUBJECT'S FIRST NAME/MI: KENNETH/W
 O.L. #: 130116
 DEPT/AGENCY: MGT/1
 CASE/REPORT: T
 TEST TYPE: T
 ALCO TARGET VALUE: 078 ✓
 ALCO S/N: 77111 ✓

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.70 in		
ALCO TARGET	.077	15:44
BLANK TEST	.000	15:45
INTERNAL STANDARD	VERIFIED	15:45
ALCO TV 29.70 in	.075 ✓	15:45
BLANK TEST	.000	15:46
SUBJECT SAMPLE	.000 <i>CSB 11/4/08</i>	15:47
BLANK TEST	.000	15:48
ALCO TV 29.70 in	.075	15:48
BLANK TEST	.000	15:49

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130116

OCTOBER 24, 2008 ✓
TIME 15:40

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
 PROGRAM: OKAY
 SOFTWARE DATE: 02/20/01
 HEATERS
 SAMPLE CHAMBER: 49c
 BREATH TUBE: 42c
 BAROMETER: 29.70 in
 FLOW DETECTOR: OKAY
 PUMP
 HIGH SPEED: OKAY
 DETECTOR: OKAY
 FILTERS: OKAY
 QUARTZ STANDARD: OKAY
 CALIBRATION: OKAY

PRINTER TEST

```
! " # $ % & ' ( ) * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢
```