

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JAN 25 08 2010

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130116 ✓

Name Kenneth Parker ID# 5578 Date 12-31-09 ✓

Agency McGrath Public Safety Phone # 524-3075

Instrument Location McGrath AST ✓

---

Alco S/N 7111 ✓ Target Value .078 High Pressure 700

Alco Test Values .074 ✓ .073 ✓

1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature [Signature] ✓

(OVER) [Signature] 1/9/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this \_\_\_\_\_ day of \_\_\_\_\_, 2010.

(Notary Seal Stamp)

Carolyn M. Noland  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm breath Test Instrument  
State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N 130116 ✓

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130116 ✓

DECEMBER 31, 2009

*COS 1/9/10*

OPERATOR'S NAME:  
PARKER/KENNETH/W  
OPERATOR'S NUMBER: 5578  
SUBJECT'S LAST NAME:  
PARKER  
SUBJECT'S FIRST NAME/MI :  
KENNETH/W  
O.L. #: 130116  
DEPT/AGENCY: MCGI  
CASE/REPORT: T  
TEST TYPE: T  
ALCO TARGET VALUE: 078 ✓  
ALCO S/N: 77111 ✓

### --- BREATH ANALYSIS ---

.078 ADJUSTED FOR 30.01 in ✓		
ALCO TARGET	.078	09:24
BLANK TEST	.000	09:25
INTERNAL STANDARD	VERIFIED	09:25
ALCO TV 30.01 in	.074	09:26
BLANK TEST	.000	09:26
SUBJECT SAMPLE	.000	09:27
BLANK TEST	.000	09:28
ALCO TV 30.01 in	.073	09:28
BLANK TEST	.000	09:29

*COS 1/9/10*

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130116 ✓

DECEMBER 31, 2009  
TIME 09:19 ✓

### --- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	30.01 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJKLHINOPQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~■