

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

MAR 25 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130116 ✓
Name	Michael Carpenter	ID#	6714
Date	3-22-10		
A Agency	Alaska State Troopers	Phone #	524-3052
Instrument Location	Capt Snow Building		McCorath AK ✓
B Alco S/N	77111	Target Value	.078 ✓
High Pressure	700		
Alco Test Values	.073 ✓	.073 ✓	
1 st Alco		2 nd Alco	
Signature	Mich Carpen		

(OVER)

(Do Not write in the area below)

BMB
4/11/10

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 5th day of April, 2010.

Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


DataMaster cdm S/N

130116 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THIS TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST



ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130116

MARCH 22, 2010

OPERATOR'S NAME:
CARPENTER/MICHAEL/R
OPERATOR'S NUMBER: 6714
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
L

BMB
4/11/10

O.L. #: L
DEPT/AGENCY: NC60
CASE/REPORT:
TEST TYPE: D
ALCO TARGET VALUE: .078 ✓
ALCO S/N: 77111

— BREATH ANALYSIS —

.078 ADJUSTED FOR 29.88 in ✓		
ALCO TARGET	.077	14:53
BLANK TEST	.000	14:54
INTERNAL STANDARD	VERIFIED	14:54
ALCO TV 29.88 in	.073 ✓	14:55
BLANK TEST	.000	14:56
SUBJECT SAMPLE	.000	14:57
BLANK TEST	.000	14:58
ALCO TV 29.88 in	.073 ✓	14:58
BLANK TEST	.000	14:59

BMB
4/11/10



DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130116 ✓

MARCH 22, 2010 ✓
TIME 14:46

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	48c
BAROMETER:	29.88 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;=<?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~