

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130117</u>		
Supervisor/Operator Performing the Verification Procedure:		
Name <u>KYLE VALERIO</u>	ID# <u>2093</u> Date <u>4/15/2011</u>	
A Agency <u>KODIAK POLICE DEPARTMENT</u> Phone # <u>(907)486-8000</u>		
Instrument Location <u>2160 MILL BAY ROAD, KODIAK, AK. 99615</u>		
B Alco S/N <u>X172961</u> Target Value <u>.081</u> High Pressure <u>950</u>		
Alco Test Values	<u>.078</u> 1 st Alco	<u>.078</u> 2 nd Alco
Signature <u>[Handwritten Signature]</u>		
(OVER)		


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 15 day of JUNE, 2010. [Handwritten Initials]

[Handwritten Signature] (Notary Seal Stamp)
Daniel L. Nicholas Kyle Klopfer
Notary Public, State of Alaska
Commission Expires with Office



CRF
4/22/11

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130117

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130117

APRIL 15, 2011 CRF 4/22/11

OPERATOR'S NAME:
VERIFICATION OF CAL
OPERATOR'S NUMBER: 2093
SUBJECT'S LAST NAME:
TEST
SUBJECT'S FIRST NAME/MI :
TEST

D.L. #: XXXXXXX
DEPT/AGENCY: KOD1
CASE/REPORT: 201100000
TEST TYPE: U
ALCO TARGET VALUE: .081
ALCO S/N: X172961

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.19 in
ALCO TARGET .081 17:08
BLANK TEST .000 17:10
INTERVAL STANDARD VERIFIED 17:10
ALCO TU 30.17 in .078 17:10
BLANK TEST .000 17:11
SUBJECT SAMPLE .000 17:11 CRF 4/22/11
BLANK TEST .000 17:12
ALCO TU 30.17 in .078 17:13
BLANK TEST .000 17:14

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130117

APRIL 15, 2011
TIME 17:14

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.17 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcde!ghijklmno
pqrstuvwxyz{|}~