

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

SEP 08 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130117</u> ✓
Name <u>KYLE VALERIO</u>		ID# <u>2093</u> Date <u>9/4/2009</u> ✓
A	Agency <u>KODIAK POLICE DEPARTMENT</u>	Phone # <u>(907)486-8000</u>
Instrument Location <u>217 LOWER MILL BAY ROAD, KODIAK, AK, 99615</u>		✓
Alco S/N <u>X172938</u> ✓		Target Value <u>.087</u> ✓ High Pressure <u>1150</u>
B	Alco Test Values	
	<u>.083</u> 1 <sup>st</sup> Alco	<u>.084</u> 2 <sup>nd</sup> Alco
Signature 		

(OVER)

BMB  
9/15/09


(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 18<sup>th</sup> day of Sept, 2009.

  
Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130117 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130117  
SEPTEMBER 04, 2009

*BMC*  
*9/15/09*

OPERATOR'S NAME:  
VALERIO, KYLE D  
OPERATOR'S NUMBER: 2893  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI :  
NEW ALCO  
O.L. #: XXXXXX  
DEPT/AGENCY: KOD1  
CASE/REPORT: 200900000  
TEST TYPE: U  
ALCO TARGET VALUE: .087 ✓  
ALCO S/N: X172938

— BREATH ANALYSIS —

.087 ADJUSTED FOR 30.01 in ✓  
ALCO TARGET .087 11:44  
BLANK TEST .000 11:45  
INTERNAL STANDARD VERIFIED 11:45  
ALCO TU 30.01 in .083 11:46  
BLANK TEST .000 11:46  
SUBJECT SAMPLE *BMC* .000 11:47  
BLANK TEST .000 11:48  
ALCO TU 30.01 in *9/15/09* .084 11:48  
BLANK TEST .000 11:49

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130117  
SEPTEMBER 04, 2009  
TIME 11:49

— DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/28/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 43c  
BAROMETER: 30.01 in  
FLOW DETECTOR: OKAY  
PUMP  
~~HIGH SPEED: OKAY~~  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST  
!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFG  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmno  
pqrstuvwxyz{|}~■