

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

APR 13 2010

Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130117</u> ✓	
Name <u>KYLE VALERIO</u>	ID# <u>2093</u> Date <u>4/8/2010</u> ✓
A Agency <u>KODIAK POLICE DEPARTMENT</u> Phone # <u>(907)486-8000</u>	
Instrument Location <u>217 LOWER MILL BAY ROAD, KODIAK, AK. 99615</u> ✓	
B Alco S/N <u>X301573</u> ✓ Target Value <u>.084</u> ✓ High Pressure <u>1150</u>	
Alco Test Values	<u>.080</u> ✓ <u>.080</u> ✓ 1 <sup>st</sup> Alco                      2 <sup>nd</sup> Alco
Signature	
(OVER)	

*Handwritten initials and date:*  
BMS  
4/15/10


(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 26<sup>th</sup> day of April, 2010.

  
(Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N: 130117

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130117

APRIL 08, 2010

OPERATOR'S NAME:  
VALERIO, KYLE D  
OPERATOR'S NUMBER: 2093  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI :  
TEST

O.L. #: XXXXXX  
DEPT/AGENCY: KOD1  
CASE/REPORT: 201000000  
TEST TYPE: U  
ALCO TARGET VALUE: .084  
ALCO S/N: X301573

--- BREATH ANALYSIS ---

.084 ADJUSTED FOR 30.19 in  
ALCO TARGET .084 15:07  
BLANK TEST .000 15:08  
INTERNAL STANDARD VERIFIED 15:08  
ALCO TU 30.19 in .080 15:08  
BLANK TEST .000 15:09  
SUBJECT SAMPLE .000 15:10  
BLANK TEST .000 15:10  
ALCO TU 30.19 in .080 15:11  
BLANK TEST .000 15:12

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130117

APRIL 08, 2010  
TIME 15:12

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 43c  
BAROMETER: 30.17 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJKL\*NOPQRSTUWXYZ\^\_`abcdefg hijklmno  
pqrstuvwxyz{|}~