

# VERIFICATION OF CALIBRATION REPORT

SEP 20 2010

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		
DataMaster cdm S/N <u>130117</u> ✓		
Name <u>KYLE VALERIO</u>	ID# <u>2093</u> ✓ Date <u>9/15/2010</u> ✓	
A Agency <u>KODIAK POLICE DEPARTMENT</u> Phone # <u>(907)486-8000</u>		
Instrument Location <u>217 LOWER MILL BAY ROAD, KODIAK, AK. 99615</u>		
B Alco S/N <u>X301573</u> ✓ Target Value <u>.084</u> ✓ High Pressure <u>600</u>		
Alco Test Values	<u>.081</u> ✓ 1 <sup>st</sup> Alco	<u>.079</u> ✓ 2 <sup>nd</sup> Alco
Signature <u>[Signature]</u>		

(OVER)

(Do Not write in the area below)

*[Handwritten initials]*  
9/23/10


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 9<sup>th</sup> day of Nov, 2010.

[Signature] (Notary Seal Stamp)  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



**VERIFICATION OF CALIBRATION REPORT**    **SEP 20 2010**  
*of DataMaster cdm breath Test Instrument*  
*State of Alaska*  
**Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program**

DataMaster cdm S/N 130117

(CONTINUED FROM FRONT PAGE)

**C**            TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

**NONDRINKING SUBJECT TEST**

ALASKA DEPARTMENT OF PUBLIC SAFETY  
 DATAMASTER cdm 130117  
 SEPTEMBER 15, 2010  
 OPERATOR'S NAME: VALERIO, KYLE D  
 OPERATOR'S NUMBER: 2093  
 SUBJECT'S LAST NAME: VERIFICATION OF CAL  
 SUBJECT'S FIRST NAME/MI : TEST  
 O.L. #: XXXXXXX  
 DEPT/AGENCY: KOD1  
 CASE/REPORT: 201000000  
 TEST TYPE: U  
 ALCO TARGET VALUE: .004  
 ALCO S/N: X301573

*AMB*  
*9/23/10*

--- BREATH ANALYSIS ---

.004 ADJUSTED FOR 30.32 in  
 ALCO TARGET            .005            12:29  
 BLANK TEST              .000            12:31  
 INTERNAL STANDARD    VERIFIED    12:31  
 ALCO TV 30.32 in      .001            12:31  
 BLANK TEST              .000            12:32  
 SUBJECT SAMPLE        .000            12:32  
 BLANK TEST              .000            12:33  
 ALCO TV 30.32 in      .079            12:33  
 BLANK TEST              .000            12:35

*AMB*  
*9/23/10*

**DIAGNOSTIC CHECK**

ALASKA DEPARTMENT OF PUBLIC SAFETY  
 DATAMASTER cdm 130117  
 SEPTEMBER 15, 2010  
 TIME 12:36

--- DIAGNOSTIC CHECK ---

COMPUTER:                OKAY  
 PROGRAM:                OKAY  
 SOFTWARE DATE:        02/20/01  
 HEATERS  
 SAMPLE CHAMBER:        43c  
 BREATH TUBE:            43c  
 BAROMETER:              30.32 in  
 FLOW DETECTOR:         OKAY  
 PUMP  
 HIGH SPEED:             OKAY  
 DETECTOR:                OKAY  
 FILTERS:                 OKAY  
 QUARTZ STANDARD:      OKAY  
 CALIBRATION:            OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF  
 GHIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
 qrstuvwxyz{|}~#