

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130117

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130117

OCTOBER 25, 2010

OPERATOR'S NAME:
VALERIO, KYLE D
OPERATOR'S NUMBER: 2093
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
TEST

O.L.#: XXXXXXXX
DEPT/AGENCY: K001
CASE/REPORT: 20100000
TEST TYPE: U
ALCO TARGET VALUE: 004
ALCO S/N: X301573

--- BREATH ANALYSIS ---

.004 ADJUSTED FOR 29.60 in
ALCO TARGET .083 14:42
BLANK TEST .000 14:43
INTERNAL STANDARD VERIFIED 14:43
ALCO TU 29.60 in .077 14:43
BLANK TEST .000 14:44
SUBJECT SAMPLE .000 14:44
BLANK TEST .000 14:45
ALCO TU 29.60 in .078 14:46
BLANK TEST .000 14:46

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130117

OCTOBER 25, 2010
TIME 14:46

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 45c
BREATH TUBE: 41c
BAROMETER: 29.60 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~