


FEB 02 2010 ²⁰¹¹ _{3/9/11}

VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument
State of Alaska*

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130119</u> ✓
Name	<u>Robin P Brennan</u>	ID# <u>4542</u> ✓ Date <u>1-29-11</u> ✓
A Agency	<u>Soldotna Police Dept</u>	Phone # <u>907-262-4455</u>
Instrument Location	<u>Soldotna PD</u>	
B Alco S/N	<u>X301536</u> ✓	Target Value <u>.081</u> ✓ High Pressure <u>1000</u>
Alco Test Values	<u>.084</u> ✓ 1st Alco	<u>.085</u> ✓ 2nd Alco
Signature		


(OVER)

*BMB
2/8/11*

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 14th day of MAR, 2010 ²⁰¹¹


 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130119

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130119

JANUARY 29, 2011

OPERATOR'S NAME:

BRENNAN

OPERATOR'S NUMBER: 4542

SUBJECT'S LAST NAME:

IDENTIFICATION OF TOL

SUBJECT'S FIRST NAME/MI :

NO

C.L. #: NO

DEPT/AGENCY: SLD1

CASE/REPORT: NO

TEST TYPE: U

ALCO TARGET VALUE: .081

ALCO S/N: X0R153S

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.99 in
ALCO TARGET .081 20:09
BLANK TEST .000 20:10
INTERNAL STANDARD VERIFIED 20:11
ALCO TO 29.99 in .004 20:11
BLANK TEST .000 20:12
SUBJECT SAMPLE .000 20:12
BLANK TEST .000 20:13
ALCO TO 29.99 in .085 20:13
BLANK TEST .000 20:14

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130119

JANUARY 29, 2011

TIME 20:15

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 03/09/08

HEATERS

SAMPLE CHAMBER: 45c

BREATH TUBE: 45c

BAROMETER: 29.99 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#&'()*+,-./:0123456789:;<=>?@BCDEFG
HIJKLMNPOQRSTUVWXYZ[\]^_`abodefghijklmno
pqrstuvwxyz{|}~