


# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska


Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130119</u> ✓
Name <u>Tobin P Brennan</u>	ID# <u>4542</u>	Date <u>2-3-09</u> ✓
A Agency <u>Soldotna Police Department</u>		Phone # <u>(907) 262-4455</u>
Instrument Location <u>Soldotna Police Department (44510 Sterling Hwy, Soldotna Alaska 99669)</u> ✓		
B Alco S/N <u>X124650</u> ✓		Target Value <u>.074</u> ✓
		High Pressure <u>550</u>
Alco Test Values		
	<u>.079</u> ✓ 1 <sup>st</sup> Alco	<u>.080</u> ✓ 2 <sup>nd</sup> Alco
Signature		
COS 2/17/09		
(OVER)		

Under the Alaska Rules of Evidence, I certify that:  
(Do Not write in the area below)

- (a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Tobin P Brennan, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed



02/18/09

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130119

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130119

FEBRUARY 03, 2009

OPERATOR'S NAME: *COS 2/17/09*  
BRENNAN  
OPERATOR'S NUMBER: 4542  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/NI :  
UOC

O.L. #: UOC  
DEPT/AGENCY: SLD1  
CASE/REPORT: UOC  
TEST TYPE: U  
ALCO TARGET VALUE: .074 ✓  
ALCO S/N: X124650

### --- BREATH ANALYSIS ---

.074 ADJUSTED FOR 29.57 in  
ALCO TARGET .073 15:54  
BLANK TEST .000 15:54  
INTERNAL STANDARD VERIFIED 15:54  
ALCO TV 29.57 in .079 ✓ 15:55  
BLANK TEST .000 15:56  
SUBJECT SAMPLE *COS 2/17/09* .000 15:56  
BLANK TEST .000 15:57  
ALCO TV 29.57 in .000 ✓ 15:57  
BLANK TEST .000 15:58

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130119

FEBRUARY 03, 2009 ✓  
TIME 15:51

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 03/29/00  
HEATERS  
SAMPLE CHAMBER: 45c  
BREATH TUBE: 45c  
BAROMETER: 29.57 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

! " # % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~