

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 28 2009

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130119 ✓

Name Tobin P Brennan ID# 4542 Date 7-26-09

A Agency Soldotna Police Department Phone # (907) 262-4455

Instrument Location Soldotna Police Department (44510 Sterling Hwy, Soldotna Alaska 99669) ✓

Alco S/N X301573 ✓ Target Value .080 ✓ High Pressure 1050

B Alco Test Values .080 ✓ .080 ✓
1st Alco 2nd Alco

Signature [Handwritten Signature] ✓

COB
8/27/09

(OVER)


(Do Not write in the area below)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature: Nita J. Bolz]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 17th day of Sept, 2009.

[Handwritten Signature: Carolyn M. Noland]
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130119

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130119

JULY 26, 2009

OPERATOR'S NAME:

BRENNAN

OPERATOR'S NUMBER: 4542

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

NO

O.L. #: NO

DEPT-AGENCY: SLD1

USE REPORT: NO

TEST TYPE: V

ALCO TARGET VALUE: .000

ALCO S/N: 7281573

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.85 in

ALCO TARGET .079 06:09

BLANK TEST .000 06:09

INTERNAL STANDARD (VERIFIED) 06:09

ALCO IN 29.85 in .000 06:10

BLANK TEST .000 06:10

SUBJECT SAMPLE .000 06:11

BLANK TEST .000 06:12

ALCO IN 29.85 in .000 06:12

BLANK TEST .000 06:13

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130119

JULY 26, 2009

TIME 06:06

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 03/29/00

HEATERS

SAMPLE CHAMBER: 45c

BREATH TUBE: 45c

BAROMETER: 29.85 in

FLOW DETECTOR: OKAY

NO

LEAK CHECK: OKAY

DETECTOR: OKAY

FILTERS: OKAY

DATE STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

11/15/2009 06:06:06 (S/N: 7281573) (S/N: 7281573)
ALCO S/N: 7281573 (S/N: 7281573) (S/N: 7281573)
DATE: 07/26/09 06:06:06