

FEB 07 2011
2011
3/9/11

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure
 Name: John Pernell ID#: 6260 Date: 1-28-11 ✓
 A Agency: Fairbanks Airport Police Phone #: 474-2530
 Instrument Location: 5195 Brumbaugh Ave, Fairbanks, AK 99709
 B Alco S/N: X124620 30121 2/4/11 ✓ Target Value: .082 ✓ High Pressure: 100 psi
 Alco Test Values: .061 ✓ 1st Alco .081 ✓ 2nd Alco
 Signature: John Pernell

(OVER)

Handwritten initials and date:
BMB
2/8/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 14th day of MAR, 2011 AN

Daniel L. Nicholas
 (Notary Seal Stamp)
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



