

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N		130121 ✓
Supervisor/Operator Performing the Verification Procedure:		
Name:	<u>Kyle Carrington</u>	ID# <u>5336</u> Date: <u>3/28/2008</u>
Agency:	<u>Fairbanks Airport Police</u> ✓	Phone #: <u>(907) 474-2555</u>
Instrument Location:	<u>Fairbanks Airport Police</u>	
Alco S/N:	<u>X301686</u> ✓	Target Value: <u>.082</u> ✓ High Pressure: <u>300 PSI</u>
Alco Test Value Average	<u>.081</u> 1 <sup>st</sup> Alco	<u>.080</u> 2 <sup>nd</sup> Alco
Signature	<u>Kyle S. Carrington</u> <span style="float: right;">CSB 4/4/08</span>	
(OVER)		

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Kyle Carrington, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed

Jeanne Swartz

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

4/12/08  
Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130121

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130121

MARCH 28, 2008

OPERATOR'S NAME:

CARRINGTON, KYLE

OPERATOR'S NUMBER: 5336

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/NI :

I

O.L. #: 0

DEPT/AGENCY: FBK6

CASE/REPORT: 0

TEST TYPE: U

ALCO TARGET VALUE: .002

ALCO S/N: X301686

### — BREATH ANALYSIS —

.002 ADJUSTED FOR 29.86 in  
ALCO TARGET .001 07:41  
BLANK TEST .000 07:42  
INTERNAL STANDARD VERIFIED 07:42  
ALCO TV 29.86 in .001 07:43  
BLANK TEST .000 07:43  
SUBJECT SAMPLE .000 07:44  
BLANK TEST .000 07:45  
ALCO TV 29.86 in .000 07:45  
BLANK TEST .000 07:46

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130121

MARCH 28, 2008  
TIME 07:38

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 03/29/00

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 45c

BAROMETER: 29.86 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

### PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; ( = ) ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ ¡ ¢